



S-WA.562

CANADA

DEPARTMENT OF TRADE AND COMMERCE

STANDARDS BRANCH

OTTAWA April 12, 1965.

APPROVAL LISTING

FAIRBANKS-MORSE COUNTER PLATFORM BEAM SCALE

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein have been listed as approved devices and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Counter platform beam scales Models 41-1000 and 41-1050, manufactured by Fairbanks-Morse and Company, St. Johnsbury, Vermont, U.S.A. and Canadian Locomotive Company Limited, Kingston, Ontario.

Rating of Apparatus: Capacities of 110 lbs., 220 lbs., 310 lbs., 320 lbs., 350 lbs., 55 Kg., 100 Kg., and 105 Kg.

Applications: Weighing commodities in general trade, warehouse and industrial weighing.

Conditions: As prescribed in P.C. 6894.

Descriptions: These devices are counter or bench type platform scales equipped with a main beam and a fractional beam. The leverage system is fabricated steel of the short and long T lever design, the long lever has a steelyard connection to a bench lever which in turn is connected to the registering beam assembly.

Testing: The standard tests for a platform beam scale shall apply.

C. S. Phillips
for C. S. Phillips,
Chief, Weights and Measures Division,
Standards Branch.

A. W. MacLean
A. W. MacLean,
Director,
Standards Branch.

Reference: SL-102-71

FAIRBANKS-MORSE COUNTER PLATFORM BEAM SCALE

