



CANADA

DEPARTMENT OF TRADE AND COMMERCE
STANDARDS BRANCH

S-WA.563

OTTAWA..... May 4, 1965.

APPROVAL LISTING

FAIRBANKS-MORSE COUNTER PLATFORM DIAL SCALE

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein have been listed as approved devices and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Counter platform dial scales Models 45-1000 and 45-1050 manufactured by Fairbanks-Morse and Company, St. Johnsbury, Vermont, U.S.A. and Canadian Locomotive Company Limited, Kingston, Ontario.

Rating of Apparatus: Capacities of 50 lbs., 100 lbs., 200 lbs., 300 lbs., 20 Kg., 50 Kg. and 100 Kg.


Application: Weighing commodities in general trade, warehouse and industrial weighing.

Conditions: As prescribed in P.C. 6894.

Description: These devices are counter or bench type platform dial scales with a spring resistance head mechanism. The leverage system is fabricated steel of the short and long T lever design, the long lever has a steelyard connection to a pair of compound bench levers which are connected to the head mechanism. The range of the zero adjustment shall not exceed one per cent of the capacity of the scale, and shall be operated by a separate tool.

Testing: The standard tests for a platform dial scale shall apply.


C. S. Phillips,
Chief, Weights & Measures Division,
Standards Branch.


R. W. MacLean,
Director,
Standards Branch.

Reference: SL-102-73

FAIRBANKS-MORSE COUNTER PLATFORM DIAL SCALE

