



CANADA

DEPARTMENT OF TRADE AND COMMERCE

STANDARDS BRANCH

S-WA.568

OTTAWA, May 31, 1965

APPROVAL LISTING

BERKEL BENCH DIAL SCALE

MODEL 26011

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Bench platform dial scale with tare beam model 26011 manufactured by Berkel Products Company Ltd., 1825 Dundas Highway East, Cooksville, Ontario.

Rating of Apparatus: Capacities of 57 pounds and 63 pounds.


Application: Weighing commodities in general trade.

Conditions: As prescribed in P.C. 6894.

Description: This device is a platform dial type bench scale with a single tare beam attached to the upper main lever by means of standoffs which project through vertical slots in the left side of the base housing. The load is resisted by a double pendulum assembly located in the dial head.

The leverage system consists of two fabricated steel box type main levers. The upper lever is supported by its fulcrum bearings and the lower main lever is suspended from the upper lever by the transfer bearings.

Testing: The standard tests for a bench type platform dial scale shall apply.

  
C. S. Phillips,  
Chief, Weights & Measures Division,  
Standards Branch.

  
R. W. MacLean,  
Director,  
Standards Branch.

Ref: SL-102-608A

BERKEL BENCH DIAL SCALE  
MODEL 26011

