



CANADA

DEPARTMENT OF TRADE AND COMMERCE
STANDARDS BRANCH

S-WA. 571

OTTAWA, June 3, 1965

APPROVAL LISTING

AVERY DORMANT PLATFORM DIAL SCALE
MODEL 5205DLB

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Dormant Platform Dial Scale with Ticket Printing Device Model 5205DLB manufactured by W. & T. Avery Limited, Soho Foundry, Birmingham, England and W. & T. Avery (Canada) Ltd., 22 Jutland Road, Toronto, Ontario.


Rating of Apparatus: Capacities from 5 ton to 20 ton.


Application: Weighing commodities in trade, shipping and warehouse weighing.

Conditions: As prescribed in P.C. 6894.

Description: This device is a dormant platform dial scale equipped with a printer. The underdeck lever system consists of two cast iron "Y" type levers coupled with a cast iron transfer lever which transmits movement to the dial head via the connecting rod. Resistance to the load is achieved by a spring resistance unit consisting of a single die cast lever, spring, rack and pinion assembly mounted on a die cast frame. There are three twin ball races in the assembly.


Testing: The standard test for a platform scale shall apply.


C. S. Phillips,
Chief, Weights & Measures Division,
Standards Branch.


R. W. MacLean,
Director,
Standards Branch.

Ref: SL-102-83

AVERY DORMANT PLATFORM DIAL SCALE
MODEL 5205D1B



5205/5940

