



Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

NOTICE OF APPROVAL - AVIS D'APPROBATION

NO.

S.WA - 576
Revision 3

DATE

March 9, 1972

AVERY PORTABLE PLATFORM FLOOR SCALE - MODEL 3901 AAG

Manufacturer: W. & T. Avery Limited
Birmingham, England

Distributor: Robert Morse Corporation Ltd.
Ville St. Pierre, Quebec.

Apparatus: Model 3901 AAG Portable Platform Scale,
with counterpoise weights and beam indication.
This scale may be equipped with a secondary
beam having either one or two graduated scales and poises.
A ratio counting scoop for ratios of 100:1 or 99:1 may be
included.

Beam indications and counterpoise weights
may be in avoirdupois units or metric units or in both.
Where both systems of weight units are used, the graduated
scales must be clearly identified between avoirdupois and
metric, and the two sets of counterpoise weights must differ
in shape and colour.

Rating: Maximum capacity - 2000 pounds or metric
equivalent.


Application: The weighing of commodities in trade.

Description: This scale is equipped with two A-shaped
cast iron levers arranged in the conventional manner. Load
indication is by beam and counterpoise weights.

The scale base, the load platform and the
pillar cap are of cast iron. The pillar is of sheet steel.
The scale is equipped with wheels.

S.WA-576
Revision 3

AVERY PORTABLE PLATFORM FLOOR SCALE - MODEL 3901AAG





S.WA - 576
Revision 3
March 9, 1972

Testing: The standard tests for platform floor scales shall apply.

References: SW - 85 - A3
SL - 102 - 290

NOTE: Approval is granted under the Weights and Measures Act. Chapter 292, and Regulations thereunder (P.C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above.

(for) 
Chief
Weights and Measures Division
Standards Branch


R. W. MacLean
Director,
Standards Branch