



DEPARTMENT OF TRADE AND COMMERCE  
STANDARDS BRANCH

S-WA.587

OTTAWA December 6, 1965.

APPROVAL LISTING  
BERKEL HANGING DIAL SCALE  
MODELS 20,001 AND 20,002

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein have been listed as approved devices and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Hanging dial scale models 20,001 (front indication) and 20,002 (back and front indication). Manufactured by S.A. Des Brevets van Berkel, Brussels, Belgium and Berkel Products Company Limited, 1825 Dundas Highway, East, Cooksville, Ontario.


Rating of Apparatus: Capacities 50 lb., 56 lb., 60 lb., 100 lb., 112 lb., 125 lb., 200 lb., 224 lb., 225 lb. and 20 kg., 40 kg., 60 kg., 100 kg.


Application: Weighing commodities in general trade.

Conditions: As prescribed in P.C. 6394.

Description: This device is a hanging dial scale equipped with twin pendulums as resistance to the load. The load is transmitted to the pendulum assembly by means of a multiple leverage system consisting of two second class levers. The suspended load receiving element may be platform, scoop, oval pan or hook type.

Testing: The standard tests for a suspended dial scale shall apply.

  
J. L. Armstrong,  
Chief, Weights and Measures Division,  
Standards Branch.

  
R. W. MacLean,  
Director,  
Standards Branch.

Reference: SL-102-57

BERKEL HANGING DIAL SCALE  
MODELS 20,001 AND 20,002

