



NOTICE OF APPROVAL
AVIS D'APPROBATION

S.W.A-686
Revision 1

February 24, 1976

SEEDERER-KOHLBUSCH EQUAL ARM BALANCES

MANUFACTURED BY: Seederer-Kohlbusch Inc.,
College Point,
New York

(for) Exact Weight Scale Co. (Canada) Ltd.,
Mississauga, Ontario

APPARATUS: Equal Arm Balances, Models 75, 75G.

<u>Model 75G</u>		<u>Model 75</u>	
<u>Style No.</u>	<u>Capacity-g</u>	<u>Style No.</u>	<u>Capacity-g</u>
948	200	904	200
952	300	908	300
956	750	912	750
960	1500	916	1500
962	3000	918	3000
963	6000	919	6000
965	9000		

RATING: Capacities are as listed above.

APPLICATION: The weighing of commodities in trade.

DESCRIPTION: These balances are of the style commonly used for laboratory work and other precision weighing applications. Provision is made for the customary adjustments of sensitivity, level and balance, and for lifting the beam.
The knife edges are of hardened steel and operate on agate bearings.
Balances having capacities of 200g and 300g have beams of aluminum, with brass beams being used on balances of higher capacity.
The column and the bows are of polished brass. The pans may be of aluminum or of brass, either plain or polished or bright plated.
All balances are mounted on a wooden base fitted with a drawer. The 75G series is enclosed in a glass housing.


February 24, 1976

.../2

TESTING:: The standard tests for equal arm balances
shall apply.

REFERENCE: GL 1153-57/E81-147

CONDITIONS OF APPROVAL: Approval is granted under the Weights
and Measures Act. S.C. 1970-71-72, Chapter 36, and the
Weights and Measures Regulations P.C. 1974-1461 of June 27,
1974 for use in Canada under the general conditions of the
said Regulations, and under any special conditions listed
above.


R. W. MacLean
Director General
Consumer Standards Directorate

S.WA-686 (Rev. 1)
February 24, 1976

SEEDERER-KOHLBUSCH EQUAL ARM BALANCES

