



Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

NOTICE OF APPROVAL

S.WA-691

OTTAWA January 13, 1970.

Approved: INTERNATIONAL BUTTER TESTING SCALE-STEELYARD TYPE - MODEL BS101

manufactured by International Scale and Equipment Ltd., Belmont, Ontario.

Apparatus Listed: Butter Testing Scale, Model BS101.

Load Rating: Capacity - 4 pounds.

Application: Weight control of dairy products.


Description: This scale consists of a steelyard having a graduated brass beam and two counterpoise weights. Graduations and numerals are indented and painted and are provided on both sides of the beam. The beam is notched for poise retention.

With the exception of bearings and pivots, the scale is constructed of brass or bronze materials. The poise and beam butt are of cast construction. The platter and support are fabricated from sheet brass. Counterpoise weights are machine-turned and are provided with adjusting holes.


Testing: The standard tests for beam scales shall apply.

References: SL-102-160
SW-85-I6

Note: Approval is granted under the Weights and Measures Act, Chapter 292, and Regulations thereunder (P.C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above.



(for) Chief,
Weights and Measures Division,
Standards Branch.


R. W. MacLean,
Director,
Standards Branch.

