


Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

## NOTICE OF APPROVAL

S.WA-707

OTTAWA April 3, 1970

Approved: HOWE RICHARDSON AIRCRAFT SCALE - MODEL HRS-100

manufactured by Robert Morse Corporation Ltd., Dominion Industrial Bldg.,  
Ville St. Pierre, Quebec.

Apparatus Listed: Aircraft Beam Scale, Model HRS-100

Load Rating: Maximum capacity - 25,000 pounds

Application: Wheel load weighing of aircraft

Conditions: The approval of these scales is restricted to aircraft weighing service.


Description: These scales are equipped with two L-shaped torsion levers and a transfer lever. Levers are of welded construction, utilizing steel pipe and structural sections.


Beam and poise indication is provided, utilizing a full capacity beam and a fractional beam. Beams are enclosed in sheet metal housings.

Scales are equipped with wheels and a towing handle. During weighing operations the wheels are retracted, permitting the scale base to rest directly on the floor.

References: SW-85-H11  
SL-102-210

Note: Approval is granted under the Weights and Measures Act, Chapter 292, and Regulations thereunder (P.C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above.

  
(for) Chief,  
Weights & Measures Division,  
Standards Branch.

  
R. W. MacLean,  
Director,  
Standards Branch.

