




Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

NOTICE OF APPROVAL

S.WA-714

OTTAWA June 2, 1970.

Approved: AVERY PLATFORM BEAM SCALE - MODELS 4219ABT, 4219ABA

manufactured by W. & T. Avery Ltd., Birmingham, England
Canadian Distributor: Robert Morse Corporation Ltd., Dominion Industrial Building,
Ville St. Pierre, Quebec.

Apparatus Listed: Platform Scales with beam indication -
Model 4219ABT - Beam mounted below platform level.
Model 4219ABA - Beam mounted on pedestal above platform level.
Beams are equipped with a capacity bar and a fractional bar with sliding poises. A
tare bar is optional.

Load Rating: Maximum capacity - 10,500 pounds
- 5,200 kilograms

Application: The weighing of commodities in trade.


Description: These scales are equipped with two L-shaped main levers and a transfer
lever, all of cast iron construction. Load transfer to the lever system is through
ball and socket bearing assemblies.


The scale housing and platform are of sheet steel with enamel finish.
These scales may be equipped with wheels for portable use.

Testing: The standard tests for a platform scale shall apply.

References: SW-85-A3 SL-102-290


Note: Approval is granted under the Weights and Measures Act, Chapter 292, and
Regulations thereunder (P.C. 6894) for use in Canada under the general
conditions of P.C. 6894, and under any special conditions listed above.


(for) Chief,
Weights and Measures Division,
Standards Branch.



R. W. MacLean,
Director,
Standards Branch.

AVERY PLATFORM BEAM SCALE - MODELS 4219ABT, 4219ABA

S.WA-714



4219ABT



4219ABA