




Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

NOTICE OF APPROVAL

S.WA-761

OTTAWA, September 29, 1969.

Approved: HOBART COUNTER PLATFORM DIAL SCALE MODELS 45-1000 and 45-1050

manufactured for Hobart Manufacturing Company Limited, Don Mills, Ontario, Canada, by Fairbanks Morse Inc., St. Johnsbury, Vermont, U.S.A., and Fairbanks Morse Canada Limited, Dorval, Quebec.

Apparatus Listed: Counter platform dial scale models 45-1000 and 45-1050.

Rating of Apparatus: Maximum capacity 300 pounds or 100 kilograms.

Application: Weighing of commodities in general trade.


Description: These devices are counter or bench type platform dial scales with a spring resistance head mechanism. The leverage system is fabricated steel of the short and long T lever design, the long lever has a steelyard connection to a pair of compound bench levers which are connected to the head mechanism.


The range of the zero adjustment shall not exceed one per cent of the capacity of the scale, and shall be operated by a separate tool.

Testing: The standard tests for a platform dial scale shall apply.

Reference: SW-85-H15
SW-85-F11
SL-102-51

Note: Approval is granted under the Weights and Measures Act, Chapter 292, and Regulations thereunder (P.C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above.


(for) Chief,
Weights and Measures Division,
Standards Branch.


R. W. MacLean,
Director,
Standards Branch.

HOBART COUNTER PLATFORM DIAL SCALE MODELS 45-1000 and 45-1050

S.WA-761

