




Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

NOTICE OF APPROVAL

S.WA-762

OTTAWA October 23, 1969.

Approved: ZHUNSHENG BENCH PLATFORM SCALE - MODEL SP901

manufactured by Zhunsheng Scales, Canton, China for Dover Pad Limited, 281 Common St., Montreal, Quebec.

Apparatus Listed: Bench Platform Beam Scale, Model SP901, with beam graduated on both sides. Optional graduation arrangements are as follows -

Avoirdupois both sides
Metric both sides
Avoirdupois one side, Metric other side.

Scales are furnished with counterpoise weights, both metric and avoirdupois weights being supplied where beams are graduated in both units.

Rating of Apparatus: Maximum capacity -

Avoirdupois - 100 pounds
Metric - 50 kilograms.

Special Condition: For scales equipped for both avoirdupois and metric weighing, every metric counterpoise weight shall be marked with a circular dot in green colour. This marking shall be of a permanent nature with the coloured dot being not less than 1/4 inch in diameter.


Application: The weighing of commodities in trade.

Description: This scale utilizes two "A" type cast iron levers, arranged in conventional manner. Registration is by means of a beam and counterpoise weights.

The scale base, platform, pillar and pillar cap are of cast iron construction with enamel finish.


The beam, poise and counterpoise weights are of brass.


Testing: The standard tests for platform beam scales shall apply.



References: SW-85-21
SL-102-283

Note: Approval is granted under the Weights and Measures Act, Chapter 292, and Regulations thereunder (P.C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above.


(for) Chief,
Weights and Measures Division,
Standards Branch.


R. W. MacLean,
Director,
Standards Branch.