



**NOTICE OF APPROVAL
AVIS D'APPROBATION**

S.WA-762

Revision 1

This circular supersedes S.WA-762
dated Oct. 23, 1969.

Ottawa, January 31, 1979

ZHUNSHENG - BENCH PLATFORM SCALES

MANUFACTURER: Zhunsheng Scales
China National Light Industry
Canton, China
Distributor -
Dover Pad Limited
Montreal, Quebec

DEVICES APPROVED: Bench platform beam scales, models as listed below:

<u>Model No.</u>	<u>Maximum Capacity</u>	
	<u>Pounds</u>	<u>Kilograms</u>
SP901	100	50
SP900	200	100

APPLICATION: The weighing of commodities in trade.

DESCRIPTION: These scales utilize two "A" type cast iron levers, arranged in conventional manner. Registration is by means of a beam and counterpoise weights.

The beam is graduated on both sides with optional graduation arrangements being as follows:

- Avoirdupois both sides
- Metric both sides
- Avoirdupois one side, metric other side.

Both metric and avoirdupois counterpoise weights are furnished with a scale having the beam graduated in both units.

The scale is principally constructed of cast iron with enamel finish. The beam, poise and counterpoise weights are of brass.

These scales carry the trade name "Gold Brand."

TESTING: The standard tests for platform beam scales shall apply.

SPECIAL CONDITION: For scales equipped for both avoirdupois and metric weighing, every metric counterpoise weight shall be marked with a circular dot in green colour. This marking shall be of a permanent nature with the coloured dot being not less than 1/4 inch in diameter.

REFERENCE: G6922-D226.

CONDITIONS OF APPROVAL: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72, Chapter 36, and the Weights and Measures Regulations P.C. 1974-1461 of June 27, 1974 for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.

John Armstrong

John Armstrong
Chief, Weights and Measures
Legal Metrology Branch

