




Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

NOTICE OF APPROVAL

S.WA-767

OTTAWA October 29, 1969.

Approved: TROEMNER COSMETIC SCALE - MODEL W2

manufactured by Henry Troemner Inc., Philadelphia, Pennsylvania, for Charles of the Ritz, Montreal, Quebec.

Apparatus Listed: Beam Indicating Counter Scale, Model W2

Rating of Apparatus: Maximum capacity - 2 ounces.

Application: The weighing of cosmetic powders in trade.

Description: This scale is of the pan-over-beam type, stabilized by a parallel check link. It is fitted with hardened steel pivots and agate bearings.


The commodity holder is a bright plated metal scoop. The scale base is made from a solid block of transparent plastic material. The beam and poise are of brass, nickel plated.

Testing: The standard tests for a counter scale shall apply.

References: SW-85-T15
SL-102-181

Note: Approval is granted under the Weights and Measures Act, Chapter 292, and Regulations thereunder (P.C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above.


(for) Chief,
Weights & Measures Division,
Standards Branch.


R. W. MacLean,
Director,
Standards Branch.

