




Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

NOTICE OF APPROVAL

S.WA-795

OTTAWA April 22, 1970.

Approved: BERKEL COUNTING SCALE - MODEL PT

manufactured by S.A. Brevets Van Berkel, Brussels, Belgium and Berkel Products Co. Ltd., 1825 Dundas Highway East, Cooksville, Ontario.

Apparatus Listed: Model PT Counting Scale, with two counting ratios and over/under indicator. Counting ratio options are -

1: 9 and 1: 99

1:11 and 1:143

1:10 and 1:100

1:12 and 1:144

Load Rating: Maximum capacity - 40 pounds

Application: The counting of commodities in trade.

Description: This scale is equipped with a single load lever, A-shaped and of cast construction. The load pan is stabilized by means of a link which extends upward within the head support column and has a pivoting locus within the headworks housing.


A steelyard connects the load lever to a pair of interconnected levers which carry the ratio cups. Balance is indicated by means of a pointer and a graduated chart.


The scale housings are of cast construction with enamel finish. The commodity holder is a metal scoop. Scales are equipped with levelling feet and a level indicator.

Testing: The standard tests for counting scales shall apply.

References: SW-85-B22 SL-102-322

Note: Approval is granted under the Weights and Measures Act, Chapter 292, and Regulations thereunder (P.C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above.


(for) Chief,
Weights and Measures Division,
Standards Branch.


R. W. MacLean,
Director,
Standards Branch.

BERKEL COUNTING SCALE - MODEL PT

S.WA-795

