


Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

NOTICE OF APPROVAL

S.WA-796

OTTAWA May 8, 1970.

Approved: HOWL RICHARDSON ANIMAL SCALE - MODEL HR-6700V.

manufactured by Robert Morse Corporation Ltd., Dominion Industrial Building, Ville St. Pierre, Quebec and 8250 Fraser Street, Vancouver, B.C.

Apparatus Listed: Single Animal Scale, Model HR-6700V. Scales may be equipped with any of the following types of indicators and auxiliary equipment -

- Double bar beam with proportional weights.
- Type registering beam.
- Double pendulum cabinet dial with analog indication - Magna-Weigh Model Series 32000 and 31000.
- Double pendulum mechanism, cabinet type, with projected digital indication - Magna-View Model Series 36000 and 37000.
- Single pendulum mechanism with projected digital indication - Magna-View Model Series 34000.
- Single pendulum digital indicator for beam scales - Weight-O-Graph Model Series 77
- Weight Printer - Mechanoprint Models MRL and MR2.

Load Rating: Maximum capacity - 3100 pounds.

Application: The weighing of livestock in trade.

Special Conditions: This scale must be fixed in a permanent installation when used in trade service.

Description: This scale is equipped with two L-shaped torsion type load levers of welded steel construction. The lever system is enclosed in a sheet steel housing.


Scales may be equipped with beams or self-balancing indicators as listed above.

The weighbridge is of welded construction, utilizing structural steel sections. The deck is of wood. The platform may be fitted with a cattle rack.

....contd..2..

HOWE RICHARDSON ANIMAL SCALE - MODEL HR-6700V


S.WA-796




Testing: The standard tests for a platform floor scale shall apply.

References: SW-85-H11 SL-102-123

Note: Approval is granted under the Weights and Measures Act, Chapter 292, and Regulations thereunder (P.C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above.


(for) Chief,
Weights and Measures Division,
Standards Branch.


R.W. MacLean
Director,
Standards Branch.