

Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

NOTICE OF APPROVAL - AVIS D'APPROBATION

NO.	S.W.A - 826
DATE	Jan. 14, 1972

OHAUS BEAM TYPE COUNTER SCALE

Manufacturer: Ohaus Scale Corporation
Florham Park, New Jersey, U.S.A.

Apparatus: Beam Type Counter Scale, Model 850 S

Rating: Capacity - 18 ounces or 610 grams.

Application: Weight control of dairy products.

Description: This scale is pan over beam type, utilizing a parallel link stabilizing mechanism. The scale is equipped with a magnetic damper.


The beam is fitted with three poises and three graduated scales, the main scale having coarse graduations and the other two having graduations suitable for fine reading and taring.


The scale frame and the beam are of cast aluminum construction. The load platform is a circular stainless steel plate.

Testing: The standard tests for beam type counter scales shall apply.

References: SW - 85 - 03
SL - 102 - 345

Conditions of Approval: Approval is granted under the Weights and Measures Act, Chapter 292, and Regulations thereunder (P.C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above.

(for) 
Chief
Weights and Measures Division
Standards Branch


R. W. MacLean
Director
Standards Branch

OHAUS BEAM TYPE COUNTER SCALE

