



Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

NOTICE OF APPROVAL - AVIS D'APPROBATION

NO.	S.WA - 839
DATE	April 27, 1972


HOWE RICHARDSON BENCH DIAL SCALE - MODEL 6400


- Manufacturer: Robert Morse Corporation Ltd.
Ville St. Pierre, Quebec.
- Apparatus: Bench Dial Scale, Model 6400
- Rating: Maximum capacity - 300 pounds or metric
equivalent.
- Application: The weighing of commodities in trade.
- Description: These scales are equipped with two load levers fabricated from steel plate and tubular sections and assembled by welding. Load resistance is by means of a pair of coil spring assemblies connected in parallel. The indicating pointer is actuated by means of a rack and pinion drive mechanism.
- The scale enclosure and load platform are of sheet steel with enamel finish. The scale is equipped with a level indicator and with adjustable levelling feet.
- Testing: The standard tests for platform dial scales shall apply.
- References: SW - 85 - H11
SL - 102 - 88

.../2

Conditions of Approval:

Approval is granted under the Weights and Measures Act, Chapter 282, and Regulations thereunder (P.C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above.

(for) 
Chief,
Weights and Measures Division
Standards Branch.


R.W. MacLean,
Director,
Standards Branch.

