



**NOTICE OF APPROVAL  
AVIS D'APPROBATION**

S.WA - 846

Ottawa, March 1, 1974

**APPROVED:** Howe Richardson Tank and Hopper Scales Models 30010, 30020, 30030, 30040.

**MANUFACTURED:** Howe Richardson Scale Company of Canada Ltd.  
Ville St. Pierre,  
Montreal, Quebec.

**APPARATUS LISTED:** Tank and Hopper Scales with fabricated steel levers.

<u>Model Number</u>	<u>Max. Capacity-Pounds</u>
30010	14,500
30020	70,000
30030	130,000
30040	224,000

**LEVER SYSTEM:** Two main levers, welded steel "T" or "L" shape depending upon scale application. The lever system may either be suspension type or on stands.

**INDICATION:** Mechanical Dial-Models 9100, 27000, 29000, 31000, 31000E4A, 32000, 33000, 35000, 36000, 37000, 37000E4A.  
Electronic Indicators - series 7700, 7800.  
Beams - full capacity and type registering.

**ACCESSORIES:** Unit Weights - Manual or automatic  
Beam Over and Under Indicator - Model 6102  
Projection Type Beam Indicator - Models 77, 77-W, 77-DW.  
Cut-offs, Encoders, Remote Indicators, must meet all the requirements and perform in accordance with the Weights and Measures Act and Regulations.  
Printers, Models MR-1, MR-2, CP-1 and CP-2.


APPLICATION: Weighing of free flowing solids or liquids  
in trade.

TESTING: The standard tests for hopper and tank scales  
shall apply.

REFERENCE: GL 1153-57/H186-740

CONDITIONS OF APPROVAL:

Approval is granted under the Weights and  
Measures Act, Regulations thereunder (P.C. 6894) for use in  
trade in Canada under general conditions of P.C. 6894 and  
under any special conditions listed above.

  
R. W. MacLean  
Director General  
Consumer Standards Director te

HOWE RICHARDSON TANK AND HOPPER SCALES

