



Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

NOTICE OF APPROVAL - AVIS D'APPROBATION

NO.	S.WA - 849
DATE	April 5, 1974

APPROVED: HOWE RICHARDSON BUILT-IN FLOOR SCALE
MODELS 2908, 3910, 3910, 3912, 3914.

MANUFACTURER: Howe Richardson Scale Company of Canada Ltd.
Ville St. Pierre,
Montreal Quebec.

APPARATUS
LISTED: Built-In Floor Scale


<u>MODEL NUMBER</u>	<u>CAPACITY TONS</u>	<u>PLATFORM SIZE-FEET</u>
2908	25	8 x 10
2910	25	10 x 10
3910	35	10 x 10
3912	35	12 x 10
3914	35	14 x 10

INDICATION: Mechanical Dial Heads - Models 31000E4A, 36000
37000, 37000E4A, with or without tare and
capacity beams.
Electronic Indicators - Series 7700, 7800.
Beams - full capacity or type recording.

ACCESSORIES: Unit Weights - manual or automatic
Beam Over and Under Indicator - Model 6102
Projection Type Beam Indicator, Models 77, 77-W,
77-DW.
Cut-offs, Encoders, Remote Indicators, must meet
all the requirements and perform in accordance
with the Weights and Measures Act and Regulations.
Printers - Models MR-1, MR-2, CP-1 and CP-2.

.../2

HOWE RICHARDSON BUILT-IN FLOOR SCALES



RATING OF
APPARATUS:

As listed above.

APPLICATION:

Weighing of commodities in trade.

CONDITIONS:

The commodity being weighed must be entirely supported by the weighing element.

DESCRIPTION:

This device is a straight lever built-in floor scale of the basic seven lever design. Levers are cast iron, pivots and bearings are hardened and precision ground, using special steels. Pivots are supported throughout their entire length. All bearings are self-aligning. Pivots and bearings are interchangeable. The bridge and deck is supported on bearings of the free floating ball bearing design.

TESTING:


The standard tests for built-in floor scales shall apply.

REFERENCE:

GL 1153-57/H186-89

CONDITIONS OF APPROVAL:

Approval is granted under the Weights and Measures Act R.S.C. 1970, c.W-7, and Regulations thereunder (P.C. 6894) for use in trade in Canada and under any special conditions listed above.


R.W. MacLean

Director General
Consumer Standards Directorate.