

D MCEVOY



Consumer and
Corporate Affairs Canada
Legal Metrology

Consommation
et Corporations Canada
Metrologie legale

RECORD OF APPROVAL ACTION

Approval Listing Number S.WA-883 Addendum 4

(Issued in advance of preparation of Formal Approval Notice)

OTTAWA, January 15, 1981

First Approved: January 18, 1974

Company Requesting Approval: The Hobart Manufacturing Co. Ltd.,
190 Railside Road
Don Mills, Ontario

Type of Device: Hobart electronic check stand scale for use with an
Electronic Cash Register as a Point-of-Sale Terminal utilizing an
optical projection indication system.

<u>MODEL NO.</u>	<u>RATING OR CAPACITY</u>	<u>REMARKS</u>
1500	25 lb x .01 lb OR 10 kg x .005 kg	
1500V	25 lb x .01 lb OR 10 kg x .005 kg	

Special conditions:

- (1) All installations must provide a visual indication of the weight which can be seen by both the customer and the operator from the normal position of "trading". In no instance should the weight display be greater than 45° from the customer's normal line of sight and the weight display must be locked in position.
- (2) When installed the remote digital indicator unit must not be located more than five(5) feet from the weighing unit.

1. The following is a list of electronic cash registers which are fully approved for installation with the Hobart models 1500 and 1500V scales:

<u>MANUFACTURER</u>	<u>MODEL</u>	<u>FORMER APPROVAL or SCHEDULE 1 LISTING</u>
NCR	C255	S 46
NCR	2140	N/A
NCR	2552	N/A
NCR	2135	S.WA-1082
Victor	530	N/A
DTS	440	N/A
Sweda	625	S283
Sweda	L50	N/A
National Semiconductor	T2500/T2600	S.WA-X130
National Semiconductor	T2000	S.WA-X137
National Semiconductor	T3000	S.WA-X138
IBM	3663 series	S271

The following is a list of electronic cash registers which are specially approved for installation with the Hobart models 1500 and 1500V scales if initially verified on/or before the expiry date. No additional installations are allowed after the expiry date.

<u>MANUFACTURER</u>	<u>MODEL</u>	<u>FORMER APPROVAL or SCHEDULE 1 LISTING</u>	<u>EXPIRY DATE</u>
TEC	BRC-10-CFX	S.WA-X130	Dec. 31, 1978
Jay Corp.	620	S.WA-X132	Dec. 31, 1978
Sharp Elect.	ER 6500	S.WA-X134	Dec. 31, 1978

Special Conditions: The TEC model BRC-10-CFX, Jay Corp. model 620, and Sharp Elect. model ER6500 do not have a weight signal that is free floating until entry for computation.


3. The following is a list of electronic cash registers which are temporarily approved for installation with the Hobart models 1500 and 1500V scales if initially verified on/or before the expiry date.

<u>MANUFACTURER</u>	<u>MODEL</u>	<u>FORMER APPROVAL or SCHEDULE 1 LISTING</u>	<u>EXPIRY DATE</u>
Sweda	425	S.WA-X133	August 31, 1980

SPECIAL CONDITIONS OF APPROVAL:

- (1) Associated with the National Semiconductor Corp. models T2500, T2600, T2000, T3000, there are 16 installations utilizing C1000 controllers which will not have the "free floating" facility added until the time of metric conversion. The appropriate District Managers will be notified of these installations.
- (2) The IBM model 3663 series allows a weight to be manually entered through a keying sequence, regardless of what weight is on the scale, whenever a weight is manually entered by a keyboard sequence the phrase "MANUAL WT" will be printed next to it.
- (3) The Sweda model 425 does not have a weight signal that is free floating until entry for computation. This model will be retrofitted with a new ROM on/or before March 31, 1980 to comply with the requirements of a free floating weight signal for all model 425's installed after December 31, 1978. All model 425's installed on/or before December 31, 1978 will not be required to have this retrofit carried out.

Reference No. ~~WA~~ G6922-H184



 John Armstrong, Chief
 Weights and Measures Division
 Legal Metrology Branch