



This revision supercedes SWA-885 dated
January 26, 1979. First approved
SWA-885 dated January 26, 1979

(Ottawa) November 20, 1979

NOTICE OF APPROVAL – AVIS D'APPROBATION

TRINER COMPUTING PARCEL POST SCALE

Model 550P and 550PM

Company Requesting Approval:

Triner Scale and Manufacturing
Company of Canada Limited
150 Telson Road
Markham, Ontario

Manufacturer:

Triner Scale and Manufacturing
Company of Canada Limited
150 Telson Road
Markham, Ontario

Apparatus:

Computing Parcel Post Scale
Avoirdupois model 550P
Metric model 550PM

Rating or Capacity:

Model 550P - 50 pounds x 2 ounces.
Model 550PM - 22 kilograms x 50 grams

Application:


Parcel post weighing and the determination of postal changes.

Description:

This scale is equipped with a pair of contre-connected A-shaped load levers of pressed steel construction. The platform suspension mechanism is parallel link type. Load resistance is by means of a double pendulum assembly.

The indicator chart is fan type with scales for weight and for several parcel post rate classifications. The scale base and indicator housings are of sheet steel with enamel finish. The load platform is of stainless steel.

TRINER COMPUTING PARCEL POST SCALE
MODEL 550P




Testing:

The standard tests for postal scales shall apply.

Reference:

G6922-T167-20

Conditions of Approval: Approval is granted under the Weights and Measures Act. S.C. 1970-71-72, Chapter 36, and the Weights and Measures Regulations P.C. 1974-1461 of June 27, 1974 for use in Canada under the general conditions of the said Regulations and under any special conditions listed above.



John Armstrong
Chief
Weights and Measures Division
Legal Metrology Branch

11)

)

)