



Consumer and  
Corporate Affairs

Consommation et  
corporations

Legal Metrology

Métrologie Légale

SWA-885

Approval No. d'approbation  
Revision 1

This revision supercedes SWA-885 dated  
January 26, 1979. First approved  
SWA-885 dated January 26, 1979

Ottawa November 20, 1979

## **NOTICE OF APPROVAL – AVIS D'APPROBATION**

### **TRINER COMPUTING PARCEL POST SCALE**

**Model 550P and 550PM**

#### **Company Requesting Approval:**

Triner Scale and Manufacturing  
Company of Canada Limited  
150 Telson Road  
Markham, Ontario

#### **Manufacturer:**

Triner Scale and Manufacturing  
Company of Canada Limited  
150 Telson Road  
Markham, Ontario

#### **Apparatus:**

Computing Parcel Post Scale  
Avoirdupois model 550P  
Metric model 550PM

#### **Rating or Capacity:**

Model 550P - 50 pounds x 2 ounces.  
Model 550PM - 22 kilograms x 50 grams

#### **Application:**


Parcel post weighing and the determination of postal charges.

#### **Description:**

This scale is equipped with a pair of contre-connected A-shaped load levers of pressed steel construction. The platform suspension mechanism is parallel link type. Load resistance is by means of a double pendulum assembly.

The indicator chart is fan type with scales for weight and for several parcel post rate classifications. The scale base and indicator housings are of sheet steel with enamel finish. The load platform is of stainless steel.

TRINER COMPUTING PARCEL POST SCALE  
MODEL 550P




Testing:

The standard tests for postal scales shall apply.

Reference:

G6922-T167-20

Conditions of Approval: Approval is granted under the Weights and Measures Act. S.C. 1970-71-72, Chapter 36, and the Weights and Measures Regulations P.C. 1974-1461 of June 27, 1974 for use in Canada under the general conditions of the said Regulations and under any special conditions listed above.

  
-----  
John Armstrong  
Chief  
Weights and Measures Division  
Legal Metrology Branch