



Consumer and
Corporate Affairs

Consommation et
corporations

Standards

Normes

**NOTICE OF APPROVAL
AVIS D'APPROBATION**

S.WA-890

Ottawa, February 20, 1975

EASTERN PORTABLE PLATFORM BEAM SCALES

Manufactured For: Eastern Scale Works Ltd.,
Toronto, Ontario
By - United Scales of America Division,
Howe Richardson Scale Company,
Rutland, Vermont

Apparatus: Portable Platform Beam Scales

<u>Model No.</u>	<u>Type of Beam</u>	<u>Platform Size</u>	<u>Max. Capacity</u>	
			<u>Pounds</u>	<u>Kilograms</u>
4802	Single	19" X 28"	1100	500
4803	Single	28" X 28"	2000	1000
4804	Double	19" X 28"	1100	500
4805	Double	28" X 28"	2000	1000
4806	Full Capacity	19" X 28"	1100	500

These scales may be equipped with a balance indicator and with a counting attachment with ratios of 100:1 and 10:1.

Rating: As listed above.

Application: The weighing of commodities in trade.


Description: These scales are equipped with two L-shaped load levers of welded steel construction. The levers are connected by means of two loop bearing assemblies, one on each side of the lever assembly.

Beams are of die cast alloy with brass poises. Counterpoise weights are of cast iron.

Load platforms and housings are of welded sheet steel with enamel finish. Scales are equipped with wheels.

February 20, 1975


EASTERN PORTABLE PLATFORM BEAM SCALES



Testing: The standard tests for platform Floor scales shall apply.

Reference: GL1153-57/E82-644

Conditions of Approval: Approval is granted under the Weights and Measures Act, R.S.C. 1970, C.W-7, and Regulations thereunder (P.C. 1461) for use in Canada under the general conditions of P.C. 1461, and under any special conditions listed above.


R.W. MacLean
Director-General
Standards Branch