



**NOTICE OF APPROVAL  
AVIS D'APPROBATION**

S.WA-896

Ottawa, February 18, 1975

**EASTERN BENCH TYPE BEAM SCALE-MODEL 4801**

**Manufactured For:** Eastern Scale Works Ltd.,  
Toronto, Ontario

By - United Scales of America Division,  
Howe Richardson Scale Company,  
Rutland, Vermont

**Apparatus:** Bench Type Beam Scale-Model 4801.  
These scales may be equipped with a balance indicator  
and with a counting attachment with ratios of 100:1 and 10:1.

**Rating:** Capacity, 300 pounds or 150 kilograms

**Application:** The weighing of commodities in trade.

**Description:** These scales are equipped with two L-shaped load  
levers of welded steel construction. The levers are connected  
by means of two loop bearing assemblies, one on each side of  
the lever assembly.

Beams are of die cast alloy with brass poises.


Counterpoise weights are of cast iron.

Load platforms and housings are of welded sheet  
steel with enamel finish. Scales are equipped with levelling  
feet.

**Testing:** The standard tests for platform bench scales shall apply.

**Reference:** GL1153-57/E82-644

**Conditions of Approval:** Approval is granted under the Weights  
and Measures Act, R.S.C. 1970, C.W-7, and Regulations thereunder  
(P.C. 1461) for use in Canada under the general conditions of  
P.C. 1461, and under any special conditions listed above.

  
R.W. MacLean  
Director-General  
Standards Branch

February 18, 1975

EASTERN BENCH TYPE BEAM SCALE-MODEL 4801

