



Consumer and  
Corporate Affairs

Consommation et  
corporations

Legal Metrology

Métrologie Légale

S.WA-911

Approval No. d'approbation  
Revision 1

Ottawa August 12, 1980

## NOTICE OF APPROVAL – AVIS D'APPROBATION

### SALTER POSTAL SCALE - MODEL 802

Manufactured by: Salter Industrial Measurement Co., Ltd.  
West Bromwich, Staffs, England  
for Britman Limited  
Oshawa, Ontario

Apparatus: Double Dial Postal Scale, Model 802.

Rating: Maximum capacity, 70 pounds x 2 ounces  
30 kilograms x 50 grams

Application: The weighing of mail and parcels for determination of postal charges.


Description: This scale has a spring resistant weighing mechanism which utilizes a pair of helical tension springs connected in parallel. The indicating pointer is activated by means of a rack and pinion mechanism. Front and back indication is provided.

The scale frame is of pressed steel construction. Housings are of sheet aluminum with self-coloured anodized finish. The load platter is sheet steel. The scale is equipped with a level indicator and four levelling feet.

Testing: The standard tests for postal scales shall apply.

Reference: G6922-S502-20

Conditions of Approval: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72, Chapter 36, and the Weights and Measures Regulations P.C. 1974-1461 of June 27, 1974 for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.

  
-----  
John Armstrong  
Chief, Weights and Measures  
Legal Metrology Branch

S.WA-911

October 4, 1976

SALTER POSTAL SCALE - MODEL 802

