



NOTICE OF APPROVAL  
AVIS D'APPROBATION

S.WA-931

February 13, 1976

EXACT WEIGHT COUNTING SCALE

MANUFACTURED BY:      General Counting Scales Inc.,  
                                 Cape Coral, Florida

(for)      Exact Weight Scale Co. (Canada) Ltd.,  
                                 Mississauga, Ontario

APPARATUS:              Counting Scale, Accucount Mark III

RATING:                  Load capacity, 20 pounds  
                                 Count ratios - 99 to 1 and 9 to 1  
   -100 to 1 and 10 to 1

APPLICATION:          The counting of commodities in trade.

DESCRIPTION:          On this scale the load scoop is suspended from a main lever which in turn acts upon a ratio lever from which are suspended two ratio scoops. The levers and the lever bearing stands are principally of aluminum construction.


Balance is indicated by means of a graduated over/under chart and an indicator pointer which is attached to the ratio lever. The indicator is visible through a glass window in the front face of the scale housing.

The scale housings are of sheet metal with enamel finish. The load scoop and the ratio scoops are of sheet aluminum.

TESTING:                  The standard tests for counting scales shall apply.

REFERENCE:              GL1153-57/G367

CONDITIONS OF APPROVAL: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72, Chapter 36, and the Weights and Measures Regulations P.C. 1974-1461 of June 27, 1974 for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.

  
R. W. MacLean  
Director General  
Consumer Standards Directorate

S.WA-931

February 13, 1976

EXACT WEIGHT COUNTING SCALE

