



Consumer and
Corporate Affairs Canada

Consommation et
Corporation Canada Ottawa/Hull
K1A 0C9

Consumer Standards
Directorate

Direction générale
des normes

RECORD OF APPROVAL ACTION

Approval listing Number -

S.WA-948

Revision 1

(Issued in advance of preparation of draft Approval Listing only
in case of interim Approval).

Ottawa, July 24, 1979

Company Requesting Approval - Britman Limited
391 Marwood Drive
Oshawa, Ontario

Type of Device

Apparatus:

Portable Beam Scale, Model IT-5

Rating:

Load capacity, 500 grams x 10 grams

Application:

The weighing of commodities in trade.

Description:

This scale utilizes a full capacity notched beam, graduated on both sides,
and equipped with a dog type poise.

The scale structure and the beam are principally of formed sheet steel
construction, assembled by means of spot welding and sheet metal screws. The
poise is of stainless steel, with a hardened steel dog. The load platter is of
stainless steel sheet with skirted edges.


Testing:

The standard tests for beam type counter scales shall apply.

Conditions of Approval:

Approval is granted under the Weights and Measures Act, S.C. 1970-71-72,
Chapter 36, and the Weights and Measures Regulations P.C. 1974-1461 of
June 27, 1974 for use in Canada under the general conditions of the said
Regulations, and under any special conditions listed above.

File No. G6900-B205


John Armstrong
Chief
Weights & Measures Division
Legal Metrology Branch

BALANCE A FLEAU BRITMAN - MODELE IT-5

