



APR - 8 1983

**NOTICE OF APPROVAL**

**AVIS D'APPROBATION**

THIS NOTICE OF APPROVAL SUPERSEDES NOTICE OF APPROVAL S.WA-950 revision 1 DATED October 18, 1976.

LE PRÉSENT AVIS D'APPROBATION REMPLACE L'AVIS S.WA-950 revision 1 EN DATE DU 18 octobre 1976.

Company: Howe-Richardson Canada Inc.  
217 Brunswick Blvd.,  
Pointe Claire, Quebec

Société: Howe-Richardson Canada Inc.  
217 Brunswick Blvd.,  
Pointe Claire, Quebec

Manufacturer: Howe-Richardson Canada Inc.  
Pointe Claire, Quebec

Fabricant: Howe-Richardson Canada Inc.  
Pointe Claire, Quebec

Type of Device: Electro-mechanical portable platform scales.

Type of Device: Balances électroniques mobiles à tablier.

MODEL NUMBER  
NO. de modèle

CAPACITY\*  
Capacité\*

PLATFORM SIZE  
Dimension du tablier

5401-DWM  
5402-DWM  
5413  
5414  
5415

All models 1000 lbs to  
2000 lbs.

19 in./po. x 28 in./po.  
28 in./po. x 28 in./po.  
19 in./po. x 28 in./po.  
28 in./po. x 28 in./po.  
28 in./po. x 28 in./po.

\*or metric equivalent

\*ou l'équivalent métrique

Accessories: Compatible accessories that meet the requirements and perform in accordance with the Weights and Measures Act, Regulations and Specifications may be used.

Accessoires: Peuvent être utilisés tous les accessoires compatibles qui satisfont aux exigences de la Loi sur les poids et mesures, du règlement et des spécifications y afférents.


Description: These scales are equipped with welded steel load levers, being identical to the lever systems used on models 5401 and 5402 which are approved on S.WA-660.

Description: Les balances sont équipées d'un système de leviers de charge en acier soudé, semblable à ceux que comportent les modèles 5401 et 5402 approuvés dans l'avis S.WA-660.

The lever system acts upon a strain gauge load cell with the output fed to an electronic digital weight indicator, model SSD-300.

Le système de leviers agit sur un extensomètre à signal de sortie qui alimente un indicateur de poids électronique et numérique, modèle SSD-300.

.../2



- 3 -

Description: Continued

The load platform and scale housings are of welded sheet steel with enamel finish.

Model numbers 5413, 5414 and 5415 have the load cells mounted under the platform and connected to a remote indicator.

Option: lb/kg converter switch.

Condition of Approval: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72, chapter 36, and the Weights and Measures Regulations, C.R.C.c., 1605 for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.

Reference No.: G6922-H186-18

Description: Suite

Les tabliers de charge et les carter sont en tôle d'acier soudée fini émail.

Sur les modèles 5413, 5414 et 5415, les cellules de charge sont installées sous la plate-forme et reliées à un indicateur à distance.

Options: lb/kg commutateur convertisseur.

Condition d'approbation: L'approbation est accordée conformément à la Loi sur les poids et mesures S.R.C. 1970-71-72, chapitre 36, et au règlement d'application C.R.C.c., 1605. L'emploi est autorisé au Canada sous réserve des conditions générales dudit règlement, et de toutes les conditions particulières formulées dans le présent avis.

No. de référence: H6922-H186-18

  
W.R. Virtue

Chief  
Legal Metrology Laboratories

Chef  
Laboratoires de la Métrologie légale

APR  
AVR - 8 1983