



Consumer and
Corporate Affairs Canada

Consommation et
Corporation Canada

Consumer Standards
Directorate

Direction générale
des normes

RECORD OF APPROVAL ACTION

Approval Listing Number -

S.WA-962

Revision 4

(Issued in advance of preparation of formal Approval Notice)

Ottawa, May 27, 1980

- Company Requesting Approval:
- Thermo Electric (Canada) Ltd.,
for Analogic Corp. Wakefield, Mass.
 - Metro Equipment Corporation
Sunnyvale, California
 - Masstron Scale Inc. Columbus, Ohio

Type of Device: Digital Weight Indicator

Manufactured and marketed by Analogic Corp.

Furnished to Metro Equipment Corp. and marketed under trade name "Accu-Weigh", and Masstron Scale Inc, and marketed under the name "Masstron".

Model Numbers:

<u>COMPANY</u>	<u>MODEL NUMBER</u>
Analogic Corp.	AN 5325*
Metro Equipment	MA-20K - SMA-20K
Masstron Scale Inc.	M 2000


Count Capacity: 20,000 counts x 1 or x 2 or x 5 or x 10.

All models listed above are identical except that model SMA-20K has a NEMA 4 stainless steel enclosure.

This addendum to be read in conjunction with S.WA-962 Dated October 4, 1976.

Application: These indicators may be used in conjunction with any approved load cell weighing system with which they are compatible.

File No.: G6922-A261
G6922-M506
G6922-M1027

.....

 John Buchanan
 Acting Chief
 Weights and Measures Division
 Legal Metrology Branch