



Consumer and
Corporate Affairs

Consommation et
corporations

Standards

Normes

**NOTICE OF APPROVAL
AVIS D'APPROBATION**

S.WA-969

Ottawa, _____

DETECTO DORMANT PLATFORM SCALES SERIES
2400, 2800, 2500, 2900

Manufacturer: Detecto Scales Inc.,
Brooklyn, New York

Apparatus: Dormant Platform Dial Scales

<u>Model Series</u>	<u>Base Style</u>	<u>Max. Capacity - Lb.</u>
2400	self contained	6500
2800	semi-self contained	6500
2500	self contained	20000
2900	semi-self contained	20000

Rating: Maximum capacities are as listed above.
Scales may be furnished in equivalent metric capacity.

Application: The weighing of commodities in trade.


Description: These scales are equipped with two L-shaped load levers of welded steel construction. Load bearings are link suspension type with outboard mounting arrangement.

Dial head mechanism is double pendulum type with rack and pinion indicator drive. All models are equipped with tare and capacity beams. Series 2500 and 2900 may also be furnished with unit weights.


Series 2400 and 2500 scales may be either floor-mounted or used as a built-in installation. Series 2800 and 2900 are intended for built-in installation only.

.../2

DETECTO DORMANT PLATFORM SCALES



SERIES 2400




SERIES 2500

.../2


Testing: The standard tests for platform floor scales shall apply.

Reference: G6900-D165-10

Conditions of Approval: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72, Chapter 36, and the Weights and Measures Regulations P.C. 1974-1461 of June 27, 1974 for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.



J. L. Armstrong
Chief,
Standards Laboratory



J. D. Buchanan
Acting Chief,
Weights and Measures Division

Legal Metrology and Laboratory Services