



First approved Feb. 14, 1977

Ottawa

August 12, 1980

NOTICE OF APPROVAL – AVIS D'APPROBATION

DETECTO - DORMANT PLATFORM FLOOR SCALES

MANUFACTURER: Detecto Scales Company
103-00 Foster Ave.
Brooklyn, New York

TYPE OF DEVICE: Dormant platform floor scales with mechanical dial indication, or with electronic digital readout in conjunction with a load cell.

<u>Model Series</u>		<u>Base Style</u>	<u>Max. Capacity</u>	
<u>Mech. Dial</u>	<u>Digital</u>		<u>Pounds</u>	<u>Kilograms</u>
2400	8240	Self Contained	8000	4000
2800	8280	Semi-Self Contained	8000	4000
2600	----	Self Contained	20000	10000
2700	----	Semi-Self Contained	20000	10000

Note: Series 2600 and 2700 were formerly identified as series 2500 and 2900 respectively.

APPLICATION: The weighing of commodities in trade.


DESCRIPTION: These scales are equipped with two L-shaped load levers coupled to a transfer lever

Mechanical dial scales utilize a double pendulum resistant mechanism, and are equipped with tare and capacity beams. Series 2600 and 2700 scales are also equipped with unit weights.


On scales having digital readout weight sensing is by means of a tension load cell installed at the steelyard.

FILE: G6922-D165-10

CONDITIONS OF APPROVAL: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72, Chapter 36 and the Weights and Measures Regulations (C. 1605) for use in Canada under the general conditions of the said Regulations and under any special conditions listed above.



John Armstrong
Chief, Weights and Measures
Legal Metrology Branch



READOUT - SERIES 8240, 8280

SERIES 2800