



Consumer and
Corporate Affairs

Consommation et
corporations

Legal Metrology/Métrieologie légale

D. McClellan

First Approved May 10, 1977

**NOTICE OF APPROVAL
AVIS D'APPROBATION**

S.WA-999

Ottawa, November 12, 1980

TOLEDO - SANITARY PLATFORM SCALES - MODELS 2072, 2082

Manufacturer: Toledo Scales
Division of Reliance Electric Limited
5220 Creekbank Road
Mississauga, Ontario

Type of Device: Platform counter scales with corrosion resistant features, being particularly designed for use in food processing applications.

<u>MODEL NUMBERS</u>	<u>MAXIMUM CAPACITY</u>	
	<u>Pounds</u>	<u>Kilograms</u>
2072	250	100
2082	300	130


Approved Accessories: These scales may be equipped with any approved Toledo accessories to which they are adaptable.

Application: The weighing of commodities in trade.

Description: These scales are equipped with two A-shaped load levers and a transfer lever all of cast construction and assembled in the conventional manner. The indicating head is equipped with a spring-resistant mechanism. Dials have a diameter of 20 inches. Model 2082 may be equipped with capacity and tare beams.

The platforms, cabinets and some other parts of these scales are of stainless steel, in addition to which a corrosion resistant finish has been applied to certain areas of the scale, thus permitting complete washdown of the product zone.

Both scale models have 20 in. x 20 in., load platforms.



Conditions of Approval: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72, chapter 36, and the Weights and Measures Regulations C.R.C.c., 1605 for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.

File No.: *WLE* G6922-T170-3

John Armstrong

John Armstrong, Chief
Weights and Measures Division
Legal Metrology Branch