



Consumer and  
Corporate Affairs

Consommation et  
corporations

Legal Metrology

Métrieologie Légale

S.WA-1000

Approval No. d'approbation

First Approved May 10, 1977

Ottawa, November 12, 1980

**NOTICE OF APPROVAL – AVIS D'APPROBATION**

TOLEDO - SANITARY PLATFORM SCALES - MODELS 2182, 2882

Manufacturer: Toledo Scales  
Division of Reliance Electric Limited  
5220 Creekbank Road  
Mississauga, Ontario

Type of Device: Platform counter and floor scales with corrosion resistant features, being particularly designed for use in food processing applications.

<u>MODEL NUMBER</u>	<u>PLATFORM SIZE</u> <u>Inches</u>	<u>MAXIMUM CAPACITY</u>	
		<u>Pounds</u>	<u>Kilograms</u>
2182	24 x 30	3000	1400
2882	30 x 30	3000	1400

Approved Accessories: These scales may be equipped with any approved accessories to which they are adaptable.


Application: The weighing of commodities in trade.

Description: These scales are equipped with two centre-connected load levers of "A" shape and cast construction. The indicating head is equipped with a spring-resistant mechanism. Dials have a diameter of 20 inches. Scales may also be provided with tare and capacity beams.


The platforms, cabinets, and some other parts of these scales are of stainless steel, in addition to which, a corrosion resistant finish has been applied to certain areas of the scale, thus permitting complete washdown of the product zone.

Model 2182 may be furnished either in bench mount or portable floor mount style. Model 2882 is available as a portable floor scale only.

.../2




Model 2182



Model 2882

Conditions of Approval: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72 chapter 36, and the Weights and Measures Regulations C.R.C.c., 1605 for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.

File No. *WA* G6922-T170-18

  
\_\_\_\_\_  
John Armstrong  
Chief  
Weights and Measures Division  
Legal Metrology Branch