

S. WA-1044

Approval Listing Number -

(Issued in advance of preparation of draft Approval Listing only in case of Interim Approval).

Accurate and Automatic Scale Co. Ltd.

Company Requesting Approval -

2040 Ellesmere Road
Scarborough, Ontario

Indicator

Type of Device - Accurate and Automatic Electronic Digital/Model 2001.

<u>Model No.</u>	<u>Rating or Capacity</u>	<u>Remarks</u>
2001	Approved capacity of the weight receiving element system.	

Accessories: Remote indicator and weight printer must meet all the requirements and perform in accordance with the Weights & Measures Act and Regulations.

Application: The Accurate and Automatic model 2001 Electronic Digital Indicator with one or more approved type electronic load cells which may be used with approved type load receivers which are compatible.

Note: This device is equipped with a temperature sensor and a thermostat. When the internal temperature of the device goes below 0°C the power to the device is cut off causing the display to blank out.

Date of Application for Approval -

Central File No. - G6900-A223

Lab. File No. -

.....Dec. 2, 1977.....
Date of Filing.

John Anthony
.....
Signature

Note: File Record of Approval Action with Approval Listing notices when Approval Listing number is assigned in advance of preparation of notice; file at back of book by Central File No. when Approval Listing number is not assigned.

Remove Record of Approval Action from Approval Listing Book when the company is notified of rejection, or when Special Approval Listing notice or draft of Approval Listing notice is filed with notices; replace draft of Approval Listing notice with printed notice when available.