



Consumer and
Corporate Affairs

Consommation et
corporations

Standards

Normes

RECORD OF APPROVAL ACTION

Approval Listing Number - S.WA 1050 Rev. - Mar.12/79

(Issued in advance of preparation of draft Approval Listing only
in case of Interim Approval).

Ottawa, Jan. 24, 1978

Company Requesting Approval - Howe Richardson Scale Co. of Canada Ltd.
217 Brunswick Blvd., Pointe Claire, Que.
H9R 4R7

Type of Device Howe Richardson Digital Indicator Model W.I.I.T./W.I.S.P.

Model No.

WIIT/WIST Approved capacity of the weight
receiving element.

Accessories: Remote indicator and weight printer must
meet all the requirements and perform in
accordance with the Weights & Measures
Act and Regulations.

Application: The Howe Richardson model W.I.I.T./W.I.S.T.
digital indicator with one or more approved
type electronic load cells which may be used
with approved type load receiving element
which are compatible.

Special Conditions: Devices which include the "Axle Weighing"
option must be marked "Axle Weighing is restricted for highway
weight control only." This marking must be in a location
visible to the operator; and must be of a permanent nature
with letters of 1/4" or 6 mm. minimum height in English or French.

File No. G6900-H186-33

de

John Armstrong, Chief
Weights & Measures Division
Legal Metrology Branch.

Note: File Record of Approval Action with Approval Listing notices when
Approval Listing number is assigned in advance of preparation of notice; file at
back of book by Central File No. when Approval Listing number is not assigned.

Remove Record of Approval Action from Approval Listing Book when the company is
notified of rejection, or when Special Approval Listing notice or draft of
Approval Listing notice is filed with notices; replace draft of Approval Listing
notice with printed notice when available.