



23 1981

Ottawa, April 21, 1981
le 21 avril 1981

NOTICE OF APPROVAL – AVIS D'APPROBATION

SUPERSEDES RECORD OF APPROVAL ACTION
dated March 23, 1978

REPLACE L'AVIS D'APPROBATION EN DATE DU
23 mars 1978

Manufacturer: Accurate and Automatic
Scale Companies Ltd.,
2040 Ellesmere Road
Scarborough, Ontario

Société: Accurate and Automatic Scale
Companies Ltd.,
2040 Ellesmere Road
Scarborough, Ontario

Type of Device: Low profile industrial
floor scale.

Appareil: Bascule de sol industrielle
surbaissée.

<u>MODEL NUMBER</u> <u>No. de modèle</u>	<u>MAXIMUM CAPACITY</u> <u>Capacité maximale</u>	<u>SIZE</u> <u>Dimensions</u>
FP-26-22	1000 lb OR/ou 500 kg	24" x 24"
FP-26-3030	2000 lb OR/ou 1000 kg	30" x 30"
FP-26-33	3000 lb OR/ou 1500 kg	36" x 36"
FP-26-4040	5000 lb OR/ou 2500 kg	40" x 40"
FP-26-44	10000 lb OR/ou 5000 kg	48" x 48"
FP-26-45	20000 lb OR/ou 10000 kg	48" x 60"
FP-26-46	20000 lb OR/ou 10000 kg	48" x 72"
FP-26-55	20000 lb OR/ou 10000 kg	60" x 60"
FP-26-56	20000 lb OR/ou 10000 kg	60" x 72"
FP-26-57	20000 lb OR/ou 10000 kg	60" x 84"
FP-26-66	40000 lb OR/ou 20000 kg	72" x 72"
FP-26-67	40000 lb OR/ou 20000 kg	72" x 84"
FP-26-78	40000 lb OR/ou 20000 kg	84" x 96"
FP-26-810	40000 lb OR/ou 20000 kg	96" x 120"

Approved Accessories: These devices may be used with any approved compatible accessories.

Accessoires Approuvés: Ces appareils peuvent être utilisés avec n'importe quel accessoire compatible approuvés.

Application: The weighing of industrial merchandise.

Utilisation: Pesage des marchandises industrielle.

Description: These devices are low profile electronic floor scales. The weighing element is a self contained type dormant scale arranged for top of floor mounting. The weighing element is supported by four load cells, one cell mounted in each of the four corners. Weight indication is by means of an approved compatible electronic indicator.


Description: Il s'agit de bascules de sol industrielles électroniques surbaissée. L'élément de pesage est constitué d'une bascule fixe autonome conçu pour une installation au sol; il est soutenu par quatre dynamomètres, un dans chaque coin. L'affichage pondéral est assuré par n'importe quel indicateur électronique compatible approuvé.

Conditions of Approval: Approval is granted under the Weights & Measures Act, S.C. 1970-71-72, chapter 36, and the Weights and Measures Regulations, C.R.C.c., 1605 for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.

Conditions d'approbation: L'approbation est accordée conformément à la Loi sur les poids et mesures S.R.C. 1970-71-72, chapitre 36, et au règlement d'application C.R.C.c., 1605. L'emploi est autorisé au Canada sous réserve des conditions générales dudit règlement et de toutes les conditions particulières formulées dans le présent avis.

Reference No. G6922-A223

No. de référence: G6922-A223



Mr. Robert Bruce
Acting Chief
Weights and Measures Division
Legal Metrology Branch

Chef intérimaire
Division des poids et mesures
Direction de la Métrologie légale