

RECORD OF APPROVAL ACTION

Approval Listing Number - S.WA-1062

(Issued in advance of preparation of draft Approval Listing only in case of  
Special Approval).

Company Requesting Approval - Faema Co. Ltd. for device manufactured  
1117 Davenport Rd. by Suprema S.N.C.  
Toronto, Ontario.

Type of Device Dial type bench scale - Manufactured by  
Suprema S.N.C. (Italy)

<u>Model No.</u>	<u>Rating or Capacity</u>	<u>Remarks</u>
R5	30 kg x 10 g, 25 kg x 10 g 15 kg x 5 g, 10 kg x 5 kg 5 kg x 2 g, 3 kg x 1 g	

Application: The type R5 is an automatic multi-revolution dial  
scale designed for general counter use.

These scales are not equipped with a fine zero  
adjust. If an adjustment to zero is required it  
must be done by adding or removing weight from  
an internal shot pot.

- Restrictions:
- 1) All scales must be equipped with identification  
nameplates containing the information listed  
in Part I, paragraph 21 of the regulations.
  - 2) This approval is only valid for scales having the  
notation "Min 100 g" removed or suitably covered.  
This notation is located on the scale chart  
directly below the "zero" indication.

Date of Application for Approval - Sept. 15/77

File No. G6900-S444

.....  
March 29, 1978  
Date of Filing

*L. S. ...*  
Signature

NOTE: File Record of Approval Action with Approval Listing notices when  
Approval Listing number is assigned in advance of preparation of notice; file a  
draft of book by Central File No. when Approval Listing number is not assigned.

Remove Record of Approval Action from Approval Listing Book when the company  
is notified of rejection, or when Special Approval Listing notice or draft of  
Approval Listing notice is filed with notices; replace draft of Approval Listing  
notice with printed notice when available.