

GOV'T OF CANADA

CONSUMER AND CORPORATE AFFAIRS - STANDARDS

RECORD OF APPROVAL ACTION

Approval Listing Number - S.WA-1072

(Issued in advance of preparation of draft Approval Listing only in case of Interim Approval).

Company Requesting Approval - Electroscale Corporation
15 Third Street
Santa Rosa, California 95401, U.S.A.

Type of Device Electronic bench or floor type scale incorporating strain gauge load cell(s) with an Electroscale model DR-525 digital indicator.

Model Nos. Listed on attached sheet

Capacities: Listed on attached sheet in pounds and metric equivalent.


(See page 2)

Date of Filing:

June 1, 1978

File No.

G6900-E155


John Armstrong
Chief
Weights and Measures Division
Legal Metrology Branch

NOTE: File Record of Approval Action with Approval Listing notices when Approval Listing number is assigned in advance of preparation of notice; file a back of book by Central File No. when Approval Listing number is not assigned.

Remove Record of Approval Action from Approval Listing Book when the company is notified of rejection, or when Special Approval Listing notice or draft of Approval Listing notice is filed with notices; replace draft of Approval Listing notice with printed notice when available.

Model	Platform	Scale Height	Capacity in Pounds		
LC-1212	12½" x 12½"	8"	20	50	100
LC-1818	18½" x 18½"	8"	50	100	200
LC-1824	18½" x 24½"	8"	50	100	200
LC-2424	24½" x 24½"	8"	100	200	500
LC-3030	30½" x 30½"	10"	100	200	500

