



RECORD OF APPROVAL ACTION

Approval Listing Number

S.WA-1083

revision 7

(Issued in advance of preparation of Formal Approval Notice)

Ottawa, December 4, 1980

First Approved: June 29, 1978

Company Requesting Approval: Toledo Scales  
Division of Reliance Electric Canada Ltd.  
5220 Creekbank Road  
Mississauga, Ontario

Type of Device: Checkstand scale with optional single or dual remote weight display.

<u>MODEL NO.</u>	<u>RATING OR CAPACITY</u>	<u>REMARKS</u>
8211	30 lb x .01 lb / 15 kg x .005 kg or 9.995 kg x .005 kg	Checkstand scale

261 Remote display

The scale and remote display must meet all the requirements and perform in accordance with the Weights and Measures Act and Regulations.


The following is a list of all Cash registers which are approved for use with this type of scale. Unless otherwise noted in the Expiry Date column all Registers have full approval with this type of scale. The Expiry Date is the date after which new installations will no longer be verified (initially inspected).

<u>MANUFACTURER</u>	<u>MODEL IDENTIFICATION</u>	<u>EXISTING APPROVAL OR SCHEDULE LISTING</u>	<u>EXPIRY DATE</u>
National Semiconductor	T2500/T2600	-X133	
" "	T2000	-X137	
" "	T3000	-X138	
Jay Corporation	620	-X132	Dec. 31, 1978
Sweda	425	-X133	August 31, 1980
"	625	S289	
"	L50	----	
Sharp Elect.	ER6500	-X134	Dec. 31, 1978
IBM	3663 series	S234	
NCR	C255	S41, S42	
NCR	2552		
NCR	2125	-----	
NCR	2135	S.WA-1082	
Victor	530	S226	
D.T.S.	440	S.WA-1078	
Victor	560		
NCR	2140		
D.T.S.	219, 540		

Special Conditions of Approval:

- (1) Associated with the National Semiconductor Corp. models T2500, T2600, T2000 and T3000, there are 16 installations utilizing C1000 controllers which will not have the "free floating" facility added until the time of metric conversion. The appropriate District Managers will be advised of these installations.
- (2) The IBM model 3663 series allows a weight to be manually entered through a keying sequence, regardless of what weight is on the scale, whenever a weight is manually entered by a keyboard sequence the phrase "MANUAL WT" will be printed next to it.
- (3) The Sweda 425 does not have a weight signal that is free floating until entry for computation. This model will be retrofitted with a new ROM on/or before August 31, 1980 to comply with the requirement of a free floating weight signal for all model 425's installed after December 31, 1978. All model 425's installed on/or before December 31, 1978 will not be required to have this retrofit carried out.
- (4) The NCR 2125 does not incorporate a training mode, therefore each entry must be voided and validated to prevent updating stored data. This ECR shall be interfaced with approved compatible scales that do NOT have their own display segment test mode.

Reference No. *102*, G6922-T170-23

  
\_\_\_\_\_  
John Armstrong  
Chief,  
Weights and Measures Division  
Legal Metrology Branch