



Consumer and
Corporate Affairs Canada

Consommation et
Corporation Canada

Consumer Standards
Directorate

Direction générale
des normes

RECORD OF APPROVAL ACTION

Approval Listing Number -

S.WA-1101

Revision 2

(Issued in advance of preparation of formal Approval Notice)

Ottawa, May 27, 1980

- Company Requesting Approval:
- Thermo Electric (Canada) Ltd for Analogic Corp, Wakefield, Mass.
 - Metro Equipment Corporation Sunnyvale, Ca.
 - Masstron Scale Inc. Columbus, Ohio

Type of Device: Digital Weight Indicator.

Manufactured and marketed by Analogic Corp.

Furnished to Metro Equipment Corp. and marketed under trade name "Accu-weigh", and Masstron Scale Inc., and marketed under the name of "Masstron".

Model Numbers:


<u>COMPANY</u>	<u>MODEL NUMBER</u>
Analogic Corp.	AN 5322
Metro Equipment	MB-20K - SMB-20K
Masstron Scale Inc.	M 3000

Count Capacity: 5,000, 10,000, 20,000 or 40,000 counts with multipliers up to X500.

All Models listed above are identical except that model SMB-20K has a NEMA 4 stainless steel enclosure.

Application: These indicators may be used in conjunction with any approved load cell weighing system with which they are compatible.

File No.: G6922-A261
G6922-M506
G6922-M1027

.....

 John Buchanan - Acting Chief
 Weights and Measures Division
 Legal Metrology Branch