



Consumer and
Corporate Affairs

Consommation et
corporations

Standards

Normes

RECORD OF APPROVAL ACTION

Approval Listing Number - S.WA 1112

(Issued in advance of preparation of draft Approval Listing only
in case of Interim Approval).

Ottawa, October 27, 1978

Company Requesting Approval - Superior Scale Co. Ltd.
181 Eagle Drive Dickens
P. O. Box 30, Winnipeg, Man. R3E 1T0

Type of Device

Pipe lever scale, floor mounted or suspended. May be used as a hopper scale or a general platform scale. Pipe lever construction, consisting of two pipe levers coupled via splice arms to a transverse lever. The transverse lever may be coupled to either a dial or beam or a load cell for electronic indication. The load cells used are normally 500 or 1000 pound capacities. The pipe levers vary in length from 4 ft to 10 ft. The splice arms vary in length from 2 ft to 9 ft. Capacities from 1000 pounds to 20,000 pounds or 500 kg to 10000 Kg.

(See page 2)

File No. G6922-S1024

John Armstrong
Chief
Weights and Measures Division
Legal Metrology Branch

Note: File Record of Approval Action with Approval Listing notices when Approval Listing number is assigned in advance of preparation of notice; file at back of book by Central File No. when Approval Listing number is not assigned.

Remove Record of Approval Action from Approval Listing Book when the company is notified of rejection, or when Special Approval Listing notice or draft of Approval Listing notice is filed with notices; replace draft of Approval Listing notice with printed notice when available.



Consumer and
Corporate Affairs

Consommation et
corporations

Standards

Normes

FORMULE D'ENREGISTREMENT DE L'APPROBATION

Numéro d'enregistrement d'approbation-

(Délivré avant la rédaction du projet d'enregistrement d'approbation,
seulement dans les cas d'approbation temporaire)

Ottawa, _____

Demandeur (nom de la société) -

Type d'instrument -

No. du dossier -

.....
John Armstrong
Chef, Poids et mesures
Direction de la Métrologie
légale.

REMARQUE: Classer la formule d'enregistrement de l'approbation avec les avis d'enregistrement d'approbation lorsque le numéro d'enregistrement est délivré avant la rédaction de l'avis; classer d'après le numéro du fichier central dans les cas où le numéro d'enregistrement n'est pas attribué. Enlever la formule d'enregistrement du livre d'enregistrement d'approbation lorsque la société est avisée d'un rejet ou lorsque l'avis spécial d'enregistrement d'approbation ou son projet est classé avec les avis; remplacer les projets d'avis d'enregistrement d'approbation avec les avis imprimés, lorsqu'il y a lieu.

- 2 -

Model designation

1. S/SU/10-04/04 to S/SU/10-18/10 inclusive
2. S/FM/10-04/04 to S/FM/10-18/10 inclusive

Explanation of Model Numbers - (refer to above)

S = Superior scale model

FM = Floor Mounted (tank, hopper or platform)

SU = Suspension(hopper)

10 - 10 ton capacity

04,18 = Length of platform or suspension points on splice arm
side of hopper.

04,10 = Width of platform or suspension points on pipe side of hopper.

Condition of Approval

1. This weighbridge shall be coupled to an approved compatible electronic indicator or an approved compatible dial or beam.
2. All installations of non-listed models shall be reported to this office. This is to ensure that as each new model is manufactured, it may be included in the list below for control purposes.

Approved models (to be updated on a regular basis)

<u>Model</u>	<u>Capacity</u>
S/FM/6-06/06	6000 Kg
S/FM/10-08/08	20000 lb.