



RECORD OF APPROVAL ACTION

Approval Listing Number -

S.WA-1118

(Issued in advance of preparation of draft Approval Listing only in case of Interim Approval).

Ottawa, December 7, 1978

Company Requesting Approval -

Exact Weight Scale Co. (Quebec) Ltd.
Subsidiary of Franklin Electric
1504 Notre Dame Street, West
Montreal, Quebec H3C 1L1

Type of Device

Electroweigh

"Under - Over" multiple range indicators for beam type scales.

Models M1 - Single spring - single range sensitivity and upright mounting.

- Sensitivity - $\frac{1}{2}$ oz on platform of 100:1 ratio scale
Adjustable for beam movement of 7/16" to 1 1/8"

Models M2 - Single Spring - single range horizontal mounting
Adjustable for beam movement of 7/16" to 1 1/8"

Sensitivity $\frac{1}{2}$ oz. on platform of 100:1 ratio scale.

Model 3A - Double spring multi range sensitivity - coarse range on either side of sensitive range.

Adjustable for beam movement of 7/16" to 1 1/8"

File No. G6922-E81-34

Sl.

John Armstrong
Chief

Weights and Measures Division
Legal Metrology Branch

Note: File Record of Approval Action with Approval Listing notices when Approval Listing number is assigned in advance of preparation of notice; file at back of book by Central File No. when Approval Listing number is not assigned.

Remove Record of Approval Action from Approval Listing Book when the company is notified of rejection, or when Special Approval Listing notice or draft of Approval Listing notice is filed with notices; replace draft of Approval Listing notice with printed notice when available.