



Consumer and
Corporate Affairs Canada
Legal Metrology
Branch

Consommation
et Corporations Canada
Metrologie legale

RECORD OF APPROVAL ACTION

This Approval Supercedes Schedule 1
Listing Number 278.

Approval Listing Number-

SWA-1176

(Issued in advance of preparation of formal Approval Notice)

Ottawa, June 5, 1980

Company Requesting Approval - Howe Richardson Scale Company of Canada Ltd.
217 Brunswick Blvd.
Pointe Claire, Quebec

Type of Device -

Full load cell tank scales. Floor mounted type with compression load cells of appropriate capacity and complete with rigid checking system comprising check rods and check brackets. Including steel weighbridge to accommodate the hopper or tank. The tank or hopper design must conform with the Manufacturer's recommendations and be equipped with a test weight system which meets the minimum test weight requirements recommended by Weights and Measures for hopper and tank scales.


Standard models of these tank scales include:

<u>Model</u>	<u>Maximum Capacity*</u>	<u>Load Cell Capacity</u>
FLC-30010	12 000 lb.	(4)- 5000 lb
FLC-30020	60 000 lb.	(4)-20000 lb
FLC-30030	120 000 lb.	(4)-50000 lb
FLC-30040	200 000 lb.	(4)-50000 lb
FLC-30050	400 000 lb.	(4)-100000 lb

*(Including Dead Load)

These scales may be interfaced to any approved compatible indicators.

File No ~~2015~~ G6922-H186-11



John Buchanan
Acting Chief
Weights and Measures Division
Legal Metrology Branch

