



D 115004

Legal Metrology/Métrieologie Légale

RECORD OF APPROVAL ACTION

Approval Listing Number

S.WA-3014

Addendum 1

(Issued in advance of preparation of Formal Approval Notice)

OTTAWA, March 3, 1981

This Record of Approval Action to be read in conjunction with S.WA-3014 dated December 11, 1980.

Model Number: 90-151.

The capacity of this device is changed to read:


Capacity - up to 10,000 increments.

Increment Size - .01, .02, .05, 0.1, 0.2, 0.5, 1, 2, 5, 10, 20 and 50 lbs.

- .005, .01, .02, .05, 0.1, 0.2, 0.5, 1, 2, 5, 10 and 20 kg.

The capacity of this indicator is limited to the approved capacity of the weight sensing system to which it is interfaced.

File #: *G. Mill.* G6922-F156-33



Robert Bruce
A/Chief - Weights & Measures
Legal Metrology