



H. Mcneaney

S.WA-3051

Approval No. d'approbation

Ottawa, September 28, 1981
28 septembre 1981

NOTICE OF APPROVAL – AVIS D'APPROBATION

Howe-Richardson - FULL LOAD SUSPENDED HOPPER SCALES

BASCULE A TREMIE SUSPENDUE A ACTION DIRECTE

Company: Howe-Richardson Canada Inc.
217 Brunswick Blvd.,
Pointe Claire, Quebec
H9R 4R7

Société: Howe-Richardson Canada Inc.
217, Boul. Brunswick
Pointe Claire, Quebec
H9R 4R7

Manufacturer: Howe-Richardson Canada Inc.
217 Brunswick Blvd.,
Pointe Claire, Quebec
H9R 4R7

Fabricant: Howe-Richardson Canada Inc.
217 Boul. Brunswick
Pointe Claire, Québec
H9R 4R7

Type of Device: Electronic full load cell suspended hopper scale.

Appareil: Bascule à trémie suspendue, électronique et à action directe.

<u>MODEL NO.</u> <u>No. de modèle</u>	<u>CAPACITY</u> <u>Capacité</u>	<u>No. of LOAD CELLS</u> <u>No. de dynamomètres</u>	<u>CAPACITY OF LOAD CELLS</u> <u>Cap. des dynamomètres</u>
FLC 3H10	up to 6,500 lb	3-4	3,000 lb. - 2,000 lb.
FLC 3H20	20,000 lb.	3-4	10,000 lb. - 10,000 lb.
FLC 5H30	30,000 lb.	3-4	20,000 lb. - 10,000 lb.
FLC 15H60	60,000 lb.	3-4	25,000 lb. - 20,000 lb.
FLC 30H120	120,000 lb	3-4	50,000 lb. - 50,000 lb.
FLC 50H200	200,000 lb.	3-4	100,000 lb. - 100,000 lb.

Approved Accessories: This device may be used with an approved compatible printer.

Accessoires approuvés: La présente bascule peut être utilisée avec toute imprimante compatible approuvée.

Application: The weighing of commodities in trade.

Utilisation: Pesage des marchandises dans le commerce.

Description: This device is a fully electronic suspended hopper scale. It utilizes Transducer load cell or equal mounted in tension, which are interfaced with any approved Howe-Richardson electronic digital indicators and printers.

Description: Cette bascule à trémie suspendue entièrement électronique comporte des dynamomètres de traction Transducer ou l'équivalent qui sont reliés à n'importe quel type approuvé d'indicateurs et d'imprimantes numériques électroniques de marque Howe-Richardson.

Conditions of Approval: Approval is granted under the Weights and Measures Act., S.C. 1970-71-72, chapter 36, and the Weights and Measures Regulations, C.R.C.c., 1605 for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.

Condition d'approbation: L'approbation est accordée conformément à la Loi sur les poids et mesures S.R.C. 1970-71-72, chapitre 36, et au règlement d'application C.R.C.c., 1605. L'emploi est autorisé au Canada sous réserve des conditions générales dudit règlement et de toutes les conditions particulières formulées dans le présent avis.

Reference No. G6922-H186-11

Référence No.: G6922-H186-11



Robert Bruce
Acting Chief
Weights and Measures Division
Legal Metrology Branch

Chef intérimaire
Division des poids et mesures
Direction de la Métrologie légale