



Ottawa

SEP 30 1982

**NOTICE OF APPROVAL – AVIS D'APPROBATION**

THIS NOTICE OF APPROVAL SUPERSEDES  
S.WA-T152 ISSUED September 2, 1980.

LE PRÉSENT AVIS D'APPROBATION REMPLACE  
L'AVIS S.WA-T152 ÉMIS LE 2 septembre  
1980.

Company: Industrial Scale Company Inc.  
45 East Bend Avenue, North  
Hamilton, Ontario

Société: Industrial Scale Company Inc.,  
45 East Bend Avenue, North  
Hamilton, Ontario

Manufacturer: Industrial Scale Co. Inc.  
Hamilton, Ontario

Fabricant: Industrial Scale Company Inc  
Hamilton, Ontario

Type of Device: Electronic industrial  
platform scale.

Appareil: Bascule à tablier  
industrielle et électronique.

<u>MODEL NUMBER</u> <u>No. de modèle</u>	<u>CAPACITY</u> <u>Capacité</u>	<u>PLATFORM SIZE-ft.</u> <u>Tablier en pi</u>	<u>LOAD CELL RATING</u> <u>Capacité des</u> <u>dynamometre</u> <u>LBS./en lb.</u>
VP - 04.04.06 - 2 04.06.06 - 2	2000 lb x 0.5 lb. or/ou 1000 kg x 200 gm	4 x 4 4 x 6	1000
VP - 04.04.06 - 4 04.04.03 - 4 04.06.06 - 4 04.06.03 - 4 04.08.06 - 4	4000 lb x 1 lb. or/ou 2000 kg x 500 gm	4 x 4 4 x 4 4 x 6 4 x 6 4 x 8	2000
VP - 04.06.06 - 10 04.06.03 - 10 04.08.06 - 10 04.08.03 - 10 06.08.06 - 10 06.08.03 - 10	10000 lb x 2 lb. or/ou 5000 kg x 1 kg	4 x 6 4 x 6 4 x 8 4 x 8 6 x 8 6 x 8	5000
VP - 06.08.03 - 20	20000 lb x 5 lb. or/ou 10000 kg x 2 kg.	6 x 8	10,000
VP - 04.06.08 - 40 04.08.08 - 40 06.08.08 - 40	40000 lb x 5 lb. or/ou 20000 kg x 2 kg.	4 x 6 4 x 8 6 x 8	20,000

.../2

Description: This device is a fully electronic platform scale utilizing four (4) type SB3 cantilever load cells, manufactured by Holtinger Baldwin Measurement Inc.

Weight indication is by means of approved compatible indicators.

Condition of Approval: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72, chapter 36, and the Weights and Measures Regulations C.R.C.c., 1605 for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.

Reference No.: G6922-J425

Description: Il s'agit d'une bascule à tablier entièrement électronique comportant quatre(4) dynamomètres de flexion 5B3 fabriqués par la Holtinger Baldwin Measurement Inc.

L'affichage du poids est assuré par un indicateur approuvé et compatible.

Condition d'approbation: L'approbation est accordée conformément à la Loi sur les poids et mesures S.R.C. 1970-71-72, chapitre 36, et au règlement d'application C.R.C.c., 1605. L'emploi est autorisé au Canada sous réserve des conditions générales dudit règlement, et de toutes les conditions particulières formulées dans le présent avis.

No. de référence: G6922-J425

  
W.R. Virtue

Acting Chief  
Legal Metrology Laboratories

Chef intérimaire  
Laboratoires de la Métrologie légale