



Ottawa OCT 14 1982

NOTICE OF APPROVAL – AVIS D'APPROBATION

Company: National Controls Inc.,  
2320 Airport Blvd.,  
P.O. Box 1501  
SANTA ROSA, California  
USA 95402

Société: National Controls Inc.,  
2320 Airport Blvd.,  
P.O. Box 1501  
SANTA ROSA, California  
USA 95402

Manufacturer: National Controls Inc.  
SANTA ROSA, California

Fabricant: National Controls Inc.,  
SANTA ROSA, California

Type of Device: Electronic digital  
computing counter scale.

Appareil: Balance de comptoir électro-  
nique à affichage numérique.

MODEL NUMBER  
No. de modèle

CAPACITY\*  
Capacité\*

LOAD CELL  
Dynamomètres

3200

15 x .005 kg.

MK - 12

pour avoir du poids équivalent

\*ou l'équivalent avoir du poids.

Accessories: Compatible accessories that  
meet the requirements and perform in  
accordance with the Weights and Measures  
Act, Regulations and Specifications may  
be used.

Accessoires: Tout accessoires  
compatible et conforme aux exigences de  
construction et de rendement de la Loi,  
du règlement et des prescriptions minis-  
térielles régissant les poids et  
mesures.

Description: This device is an elec-  
tronic digital computing counter scale  
with displays for weight, unit price,  
total price and annunciators for zero,  
tare and pre-pack on both the operator  
and customers sides of the scale.

Description: Il s'agit d'une balance de  
comptoir électronique à affichage numé-  
rique comportant des dispositifs  
d'affichage du poids, du prix unitaire,  
du prix total ainsi que des voyants pour  
indiquer les modes de fonctionnement  
(remise à zéro, tare et pré-emballage)  
du côté du client et du côté de  
l'opérateur.

The following controls are positioned on  
the operators side of the scale:

Les commandes suivantes sont installées  
sur la balance, du côté de l'opérateur:

1) 0-9 - keys for entering unit price or  
preset tare.

1) des touches numérotées de 0 à 9 pour  
l'introduction du prix unitaire et de  
la tare prédéterminée.