



Ottawa

NOV 25 1982

**NOTICE OF APPROVAL – AVIS D'APPROBATION**

Company: Metro Equipment Corporation  
1235 Reamwood Avenue  
Sunnyvale, California  
USA 94086

Société: Metro Equipment Corporation  
1235 avenue Reamwood  
Sunnyvale, Californie  
USA 94086

Manufacturer: Metro Equipment Corp.  
Sunnyvale, California

Fabricant: Metro Equipment Corp.  
Sunnyvale, Californie

Type of Device: Electronic digital  
weight indicator.

Appareil: Indicateur électronique à  
affichage numérique.

MODEL NUMBER  
No. de modèle

CAPACITY  
Capacité

- (a) PEP-5K
- (b) SEP-5K

5,000 COUNT BY x1, x2, x5 OR  
x10 WITH PROGRAMMABLE DECIMAL  
POINT/5 000 comptes par multiples  
de 1, 2, 5 ou 10, à point décimal  
programmable.

- (c) PEP-10K
- (d) SEP-10K

WITH 10,000 COUNT/ 10,000 comptes.

Accessories: Compatible accessories that  
meet the requirements and perform in  
accordance with the Weights and Measures  
Act, Regulations and Specifications may  
be used.

Accessoires: Peuvent être utilisés les  
accessoires compatibles qui satisfont  
aux exigences de la Loi, du règlement et  
des prescriptions ministérielles  
régissant les poids et mesures.

Description: This device is a micro-  
processor digital weight indicator that  
when interfaced with an approved compa-  
tible weight platform becomes a weighing  
system.

Description: Il s'agit d'un indicateur  
pondéral à affichage numérique commandé  
par microprocesseur qui, une fois relié  
à une plate-forme de pesage approuvée et  
compatible, constitue un ensemble de  
pesage.

The following features may be found on  
the device:

L'appareil comporte les éléments  
suivants:

- (a) 5 LED status annunciators that  
indicate Center of Zero, Net, Gross, lb,  
g, and Motion.

- (a) 5 voyants d'état à DEL qui  
indiquent le centre du zéro, le poids  
net, le poids brut, en lb ou en kg, et  
le mouvement;

Condition of Approval: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72, chapter 36, and the Weights and Measures Regulations C.R.C.c., 1605 for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.

Reference No.: G6922-M506

Condition d'approbation: L'approbation est accordée conformément à la Loi sur les poids et mesures S.R.C. 1970-71-72, chapitre 36, et au règlement d'application C.R.C.c., 1605. L'emploi est autorisé au Canada sous réserve des conditions générales dudit règlement, et de toutes les conditions particulières formulées dans le présent avis.

No. de référence: G6922-M506



W.R. Virtue

Chef  
Legal Metrology Laboratories

Chef  
Laboratoires de la Métrologie légale

NOV 25 1982