



Consumer and
Corporate Affairs Canada

Legal Metrology

Consommation
et Corporations Canada

Métrologie légale

APPROVAL No. - N° D'APPROBATION

S.WA-3145

JUL 13 1987

NOTICE OF APPROVAL

AVIS D'APPROBATION

Company: Colt Industries
Fairbanks Weighing Division
St. Johnsbury, Vermont
USA 05819

Société: Colt Industries
Fairbanks Weighing Division
St. Johnsbury, Vermont
USA 05819

Manufacturer: Colt Industries
Fairbanks Weighing Division
St. Johnsbury, Vermont

Fabricant: Colt Industries
Fairbanks Weighing Division
St. Johnsbury, Vermont

Type of Device: Electronic digital
bench scale.

Appareil: Balance de table électronique
à affichage numérique.

<u>MODEL NUMBERS</u> <u>N° de modèles</u>	<u>SIZE</u> <u>Dimensions</u>	<u>CAPACITY</u> <u>Capacité</u>
H70-4605-1	10" x 10"	10 lb x .002 lb
H70-4605-2	10" x 10"	160 oz x .05 oz
H70-4605-5	10" x 10"	5 kg x .001 kg
H70-4601-1	12" x 14"	30 lb x .01 lb
H70-4601-2	12" x 14"	60 lb x .02 lb
H70-4601-5	12" x 14"	10 kg x .005 kg
H70-4601-6	12" x 14"	25 kg x .01 kg

Description: This device is a fully
electronic stainless steel bench scale,
with the following features -
Enunciators for Center of Zero, Gross,
Net and Tare. The operator controls are:

Description: Il s'agit d'une balance de
table entièrement électronique, en acier
inoxydable, équipée de voyants qui
indique le centre du zéro (Center of
Zero), le poids brut (Gross), le poids
net (Net) et la tare (Tare) ainsi que
des commandes de fonctionnement
suivantes:

ZERO - A pushbutton switch returns the
scale to a zero balance condition.

ZERO - bouton-poussoir utilisé pour
remettre la balance à zéro;

AUTO TARE - Used to enter platter tare or
to accumulate weights for batch weighing.


AUTO TARE - sert à l'introduction d'une
tare de plateau ou à l'accumulation de
poids lors du pesage en lot;

MODE - Switches the scale to Gross or
Net.

MODE - permet de régler la balance pour
la détermination du poids brut ou du
poids net.

Special Conditions: The sealing means is
exempt from providing ready access to
her components or adjustment as per
SM3/10.

Conditions particulières: Le plombage
n'est pas tenu de permettre l'accès
facile aux autres composants ou dispo-
sitifs de réglage conformément à la
prescription ministérielle SGM3/10.



Conditions of Approval: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72, chapter 36, and the Weights and Measures Regulations C.R.C.c., 1605 for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.

Reference No.: G6922-F156-3

Conditions d'approbation: L'approbation est accordée conformément à la Loi sur les poids et mesures S.R.C. 1970-71-72, chapitre 36, et au Règlement d'application C.R.C.c., 1605. L'emploi est autorisé au Canada sous réserve des conditions générales dudit Règlement, et de toutes les conditions particulières formulées dans le présent avis.

No. de référence: G6922-F156-3

W.R. Virtue
 W.R. Virtue

Chief
 Legal Metrology Laboratories

Chef
 Laboratoires de la Métrologie légale

15 1983